

# Absolute distance measurement by dispersive interferometry using a femtosecond pulse laser

Ki-Nam Joo and Seung-Woo Kim

Department of Mechanical Engineering, Korea Advanced Institute of Science and Technology, Science Town,  
Daejeon 305-701, South Korea  
[swk@kaist.ac.kr](mailto:swk@kaist.ac.kr)

**Abstract:** We describe an interferometric method that enables to measure the optical path delay between two consecutive femtosecond laser pulses by way of dispersive interferometry. This method allows a femtosecond laser to be utilized as a source of performing absolute distance measurements to unprecedented precision over extensive ranges. Our test result demonstrates a non-ambiguity range of ~1.46 mm with a resolution of 7 nm over a maximum distance reaching ~0.89 m.

©2006 Optical Society of America

OCIS codes: 120.3180; 120.3940; 140.7090.

---

## References and links

1. M. R. Hee, J. A. Izatt, J. M. Jacobson, J. G. Fujimoto, and E.A. Swanson, "Femtosecond transillumination optical coherence tomography," *Opt. Lett.* **18**, 950-951 (1993).
2. M. Wojtkowski and A. Kowalczyk, "Full range complex spectral optical coherence tomography technique in eye imaging," *Opt. Lett.* **27**, 1415-1417 (2002).
3. K. Minoshima and H. Matsumoto, "High-accuracy measurement of 240-m distance in an optical tunnel by use of a compact femtosecond laser," *Appl. Opt.* **39**, 5512-5517 (2000).
4. C. E. Towers, D. P. Towers, D. T. Reid, W. N. MacPherson, R. R. J. Maier, and J. D. C. Jones, "Fiber interferometer for simultaneous multiwavelength phase measurement with a broadband femtosecond laser," *Opt. Lett.* **29**, 2722-2724 (2004).
5. J. Ye, "Absolute measurement of a long, arbitrary distance to less than an optical fringe," *Opt. Lett.* **29**, 1153-1155 (2004).
6. J. Ye, H. Schnatz and L. W. Hollberg, "Optical frequency combs: from precision frequency metrology to optical phase control," *IEEE J. Sel. Top. Quantum Electron.* **9**, 1041-1058 (2003).
7. S. T. Cundiff and J. Ye, "Femtosecond optical frequency combs," *Rev. Mod. Phys.* **75**, 325-342 (2003).
8. J. Schwider and L. Zhou, "Dispersive interferometric profiler," *Opt. Lett.* **19**, 995-997 (1994).
9. K. Sakai, "Michelson-type Fourier Spectrometer for the far infrared," *Appl. Opt.* **11**, 2894-2901(1972).
10. L. Lepetit, G. Chériaux and M. Joffre, "Linear techniques of phase measurement by femtosecond spectral interferometry for applications in spectroscopy," *J. Opt. Soc. Am. B* **12**, 2467-2474(1995).
11. M. Takeda, H. Ina and S. Kobayashi, "Fourier-transform method of fringe-pattern analysis for computer-based topography and interferometry," *J. Opt. Soc. Am.* **72**, 156-160(1982).
12. R. Holzwarth, T. Udem, and T. W. Hänsch, J. C. Knight, W. J. Wadsworth, and P. St. J. Russell, "Optical frequency synthesizer for precision spectroscopy," *Phys. Rev. Lett.* **85**, 2264-2267(2000).

---

## 1. Introduction

The advent of femtosecond ultrashort pulse lasers has prompted various efforts to investigate new possibilities of advanced optical interferometry that were not possible with traditional sources such as CW lasers and white light. A notable example is the exploitation of a femtosecond laser for optical coherence tomography with the aim of performing high resolution biomedical imaging by way of either low-coherence interferometry [1] or complex spectral signal processing [2]. Another is the significant extension of the measurable distance of no periodic ambiguity by means of synthetic wavelength interferometry utilizing a femtosecond laser, which was carried out in the radio-frequency domain using a sequence of higher harmonics of the repetition rate [3] or in the optical frequency domain with flexible